
Accuracy of an oscillometric blood pressure monitor during phenylephrine-induced hypertension in dogs

American Journal of Veterinary Research (AJVR) Vol 67, No. 9, September 2006, pp. 1541-1545

Conclusions & Clinical Relevance: “In normotensive dogs, oscillometric measurements of MAP and SAP agreed more closely with direct arterial pressure measurements than oscillometric measurements of DAP. Oscillometric measurement of MAP was accurate during both normotension and hypertension in dogs.”