



Premier Mobile Electric Lift Table Setup Manual



Table of Contents

Installation	2
Charging the Batteries	2
Optional Scale Features/Operation	3
Troubleshooting	4
Care & Cleaning	4
Product Warranty	5

Installation

Open packages carefully. Inspect the shipping cartons and contents for any visible damage.

The Premier Mobile Electric Lift Table arrives pre-assembled, requiring only minor attention to be fully serviceable. After removing the table from its packaging, place table in the desired position. Use the hand control to raise or lower the height of the table.

It may be necessary to lift the Premier Mobile Electric Lift Table to remove it from the pallet. Pick up the table by the stainless steel top. Lift by grasping the frame of the table base. Stainless steel base covers are attached with Velcro® tabs. Be sure to reach under edges of the base to lift by the frame, not the cover.

The Premier Mobile Electric Lift Table is protected during shipping with a white plastic film covering the stainless steel. Peel off the plastic film before using. Refer to “Care and Cleaning” instructions.

Charging the Batteries



Equipment Alert

Keep the table plugged in and in charging mode when not in use.

The Premier Mobile Electric Lift Table utilizes rechargeable batteries to enable raising and lowering the table without having to be plugged into a power outlet at all times. Over draining these batteries will inhibit their ability to fully recharge, so it is very important to keep the batteries charged. Keep the table plugged in and in charging mode when not in use.

To Charge Batteries:

To charge batteries, plug the table into a standard power outlet. Charge the table every night, overnight, and more frequently if the table is being used continuously throughout the day without being plugged in. The table needs to be unplugged for normal use.



Conversion Kit to Jumbo power pack
203-2180-01
Battery Pack
741-0012-00
Charging Cable
718-1802-00



Control with Charger
741-0010-00
Handset
403-5640-28

Bracket
750-0010-00

Scale Features/Operation

Using The Control Panel:



Push to turn unit on. Display shows the following:

AAAA – self test

Midmark– checks display segments

u 1.7 – software revision

b xx.x – battery voltage (varies with AC plug in and internal 9v battery.) A reading of 6.0v indicates the need for battery replacement; the unit will shut off at a reading of 5.5v.

0.0 – ready to weigh



Push and hold for 4 seconds until display indicates unit shift.

Display will show “conv” until the conversion is complete.



Push the RECALL button –

The display will show the message “rECL” then alternately flash between the last stable weight and the “rECL” message. Press ON/ZERO button to return to normal weighing. Works only with hold function active.

Features:

Tare Function:

Any weight within the scale capacity can be zeroed out by pressing the ON/ZERO button. When the weight is removed, the display flashes the tared weight. Press ON/ZERO button to return to normal weighing.

Weighing Function:

Display will blank out until a close range of the weight is reached. With the hold function active, the display will show the stable weight and the “hold” indication. The displayed weight will lock until the weight is removed.

Weight Hold:

Press and hold the ON/ZERO button and the RECALL button. Display will show “zero” then “h on” or “h off”. Release buttons when selection is made. Default setting is “h on”.

Automatic Shut Off:

In AC mode, the indicator will shut off in one hour. In battery mode, the indicator will shut off in one minute.

Calibration:



- 1) Install fresh 9 volt battery in indicator.
- 2) Unplug indicator's AC adapter from back of indicator.
- 3) Press and hold ON/ZERO button and LG/KG button. The display will change from “zero” to “CAL0” in about 3 seconds.
- 4) With an empty platform press the ON/ZERO button. The indicator will count down from “16” to “1” and display “C50”.
- 5) Place a 50 lb. weight on platform and press the ON/ZERO button. The indicator will countdown from “16” to “1”. Display will show “50”.
- 6) Remove the 50 lb. weight and press the ON/ZERO button. The scale is calibrated and automatically returned to normal weight mode.
- 7) Re-plug AC adapter into indicator.

Switching from AC to DC:

Unplug the AC power cord from the display. Press the ON/ZERO button to turn the indicator on.

Switching from DC to AC:

Plug the AC adapter into a standard wall socket and the cord into the AC outlet on the back of the display. Press the ON/ZERO button to turn indicator on.

Troubleshooting

Indicators:

Messages on the readout indicate operating condition:

OL Indicates weight on scale exceeds capacity.

OPEN Indicates malfunctioning of cable, or the scale needs to be calibrated.



Indicates battery is discharged.

In the event that a problem occurs, a Midmark technician is ready to assist you. Call Midmark for assistance before attempting any repairs.

In The Event Of A Problem:

In the event that a problem occurs, Midmark Technical Support is ready to assist you. In order to save time, please check the following items before calling:

- Does the readout display have a message that indicates what the problem is? Refer to “Indicators” at left for descriptions.
- Check all cables/connections.
- Unplug AC power and battery, then plug-in again to reset the scale.

Midmark Technical Support:

1-800-643-6275

Care & Cleaning

The Premier Mobile Electric Lift Table may be cleaned with any commercially available product recommended for use on stainless steel. Frequent cleaning will prolong life and help maintain the natural luster of the stainless steel.

To prevent scratching, use a soft cloth to clean top. If scrubbing is necessary, use a nylon or plastic type scrubber. When cleaning, always rub in the direction of the “grain” of the metal.

To remove minor scratches, we recommend using a 3M ScotchBrite® scouring pad rubbed lightly in the direction of the “grain” of the metal.

Never use steel wool or steel brushes when cleaning your surgery table. Steel particles may become embedded into the stainless steel and cause rust.

Avoid use of chlorine bleach for cleaning. Chlorides, bromides, iodides and thiocyanates will cause pitting, corrosion and discoloration of any stainless steel. If using bleach, clean thoroughly with water to completely remove bleach residue.

Avoid excessive moisture.

The Premier Mobile Electric Lift Table contains electronic components that may be damaged by excessive exposure to moisture. When cleaning or mopping the floor, do not leave the table in standing water. For thorough cleaning, remove the drainage trough and reservoir and wash separately. Refer to “Features/Operation” for removal instructions.

Product Warranty

This product is warranted for 1 year from the date of delivery to the original purchaser only. If equipped with an Midmark Scale, the Scale is warranted for 2 years. Warranty is only applicable if customer complies with all instructions and specifications furnished by Midmark relating to installation, care, and application. Customer agrees that they will not modify, misapply, or misuse product in any manner which deviates from the Midmark instructions. Any repairs, alterations, or services provided by parties other than Midmark or its authorized representatives may void warranty. The buyer is responsible for all freight charges on returned items.

Product Satisfaction Policy

In the event that the customer is not fully satisfied with the products purchased, Midmark may, at its own discretion, arrange for the customer's account to be credited (excluding shipping charges) or replace the product. The customer must notify Midmark of any claim of nonconformity or defect in writing within 30 days from the date of receipt. The product must be returned to Midmark in its original packaging, undamaged and showing no unusual wear within 20 days of the customer's written notice.

To return an item, follow these simple instructions:

- Call Midmark for complete instructions (800-643-6275)
- A sales representative will give you a Return Authorization Number
- If a Restocking Fee applies, you will be informed at this time
- Label the boxes with this Return Authorization Number
(All items must have a Return Authorization Number)
- Return the items, postage/freight prepaid

If necessary, the Midmark sales representative will assist you with freight arrangements.

Note: Custom orders and/or custom colors may not be returned.

Midmark Corporation - Carthage Facilities

1000 Civil War Road
Area 5
Carthage, MO 64836
1-800-643-6275

